



## CLEVELAND FLASH AND BURNING POINT APPARATUS

Standard: EN 22592, ASTM D92, AASHTO T48, IP 36/67, UNE 7075, NFT60-118, ISO 2592

### Determination of Flash & Fire Point

The apparatus is applicable for determining measurement of flash points and burning points of petroleum products (the Cleveland open cup method) except petroleum products with flash points and burning points lower than 79°C.

### Technical Specifications :

<b>Model Number</b>	<b>NL 2039 X / 002</b>
Ignition	Automatic
Heating Power of Furnace	400 W
Temperature Adjustment	Controlled Silicon (Continuous)
Dimension (mm)	500 (L) x 400 (W) x 300 (H)
Approx. Weight	6 kg
Power	220 V, 1 ph, 50 / 60 Hz

### Unit Consists Of :

Model Number	Parts Description	Qty
NL 2039 X / 002 – P 001	Thermometer IP 28C (ASTM 11C), Range -6 +400°C	1no.
NL 2039 X / 002 – P 002	Brass Cup With Handle	1no.
NL 2039 X / 002 – P 003	Flame Gas Device Unit c/w Gas Hose	1no.

\*1 Copy of Manual Instruction

### Optional Accessories :

Model Number	Accessories Description
NL 2039 X / 002 – A 001	Flame Gas Regulator