
FIGURE GATE & RUCKCivil Engineering Testing Equipment

www.nl-test.com





Opened Sound Proof Cabinet



Closed Sound Proof Cabinet

LOS ANGELES RESISTANCE ABRASION APPARATUS

Standard: ASTM C131, EN 1097-2

Determination of the Resistance to Fragmantation

Upgraded version to ECO-SMARTZ system for more stable & better performance. This apparatus used to determine the resistance of aggregates to abrasion. The drum is rotated by a speed reducer driven by an electric motor, equipped with an automatic counter. Abrasive need to charge order separately.

Technical Specifications:

Model Number	NL 1008 X / 003
Drum Speed	31 - 33 RPM
Dimension (mm)	1070 (L) x 700 (W) x 910 (H)
Approx. Weight	250 kg
Power	220 - 240 V, 50 / 60 Hz, 9.5 A, 350 W

Unit Consists Of:

Model Number	Parts Description	Qty
NL 1008 X / 003 – P 001	Control Panel	1no.
NL 1008 X / 003 – P 002	Receiving Tray	1no.

^{*1} Copy of Manual Instruction

Optional Accessories:

Model Number	Accessories Description	
NL 1008 X / 003 – A 001	Safety & Sound Proof Cabinet	
NL 1008 X / 001 – A 002	ASTM, AASHTO, UNI, CNR Abrasive	
	Charges (Set of 12 nos.)	
NL 1008 X / 001 – A 003	EN Standard Abrasive Charges	
	(Set of 12 nos.)	

Copyright 2016 NL Scientific Instruments Sdn. Bhd. All Rights Reserved. Revision 3(All information in this catalogue only for customer's reference and not for quotation. Design and dimension maybe change without prior notice.)