## BITUMEN & ASPHALT

**Civil Engineering Testing Equipment** 

www.nl-test.com







MARSHALL WATER BATH Standard: ASTM D1559, BS 598-107, AASHTO T245, DIN 1996

## Water Baths for Marshall Specimens

Used to maintain compacted asphalt specimens at constant temperature. The unit fitted with electronic temperature controller with digital display. Constructed with stainless steel inside chamber and exterior case in powder quoted painted steel sheet.

## **Technical Specifications:**

Model Number	NL 2008 X / 002
Temperature Range	From Ambient to +99.9°C
Temperature Accuracy / Resolution	± 0.5°C / 0.1°C
Usage Capacity	26 Litres
Number of Marshall Specimens	4"/100 mm - Approx. 18 pcs
	6"/150 mm - Approx. 8 pcs
Bath Dimension (mm)	600 (L) × 300 (W) × 200 (H)
Dimension (mm)	770 (L) × 370 (W) × 300 (H)
Approx. Weight	17 kg
Power	220 ~ 240 V, 2500 W, 9.5 A, 1 Ph, 50 / 60 Hz

<sup>\*1</sup> Copy of Manual Instruction

Copyright 2025 NL Scientific Instruments Sdn. Bhd. All Rights Reserved. Revision 1 (All information in this catalogue only for customer's reference and not for quotation. Design and dimension maybe change without prior notice.)