## BITLIMEN & PSPHALT Civil Engineering Testing Equipment

www.nl-test.com



# MOTORISED UNIVERSAL VERTICAL EXTRUDER

#### Remove Specimens

Used for a smooth and rapid extrusion of soil and asphalt specimen with minimum disturbance. The unit can extrude for CBR and Marshall sample. The extruded sample is held in place by a receiving table adjustable in height and easily lowered along side of the machine. The Motorized Vertical Extruder come complete with adaptor to extrude sample having diameter 100 mm and 150 mm.

## **Technical Specifications:**

Model Number	NL 2033 X / 001
Dimension (mm)	430 (L) x 640 (W) x 960 (H)
Extruder Speed	70 mm / minute
Reverse Speed	140 mm / minute
Maximum Extrude Length	230 mm
Power	380 415 V, 3 ph, 50/60 Hz,150 Watt,1Hp
Approx. Weight	130 kg

## Unit Consists Of:

Model Number	Parts Description	Qty
NL 2033 X / 001 – P 003	4" Plunger	1no
NL 2033 X / 001 – P 004	6" Mould Retaining Disc	1no
NL 2033 X / 001 – P 005	4" Mould Retaining Disc	1no

<sup>\*1</sup> Copy of Manual Instruction

Copyright 2024 NL Scientific Instruments Sdn. Bhd. All Rights Reserved. Revision 1 (All information in this catalogue only for customer's reference and not for quotation. Design and dimension maybe change without prior notice.)