

## BLAINE AIR PERMEABILITY APPARATUS

Standard: UNE 80-122, EN 196-6, ASTM C204, BS 4359

### Fineness Of Cement

The apparatus use to determine the specific surface of cement.

### Technical Specifications :

<b>Model Number</b>	<b>NL 3000 X / 001</b>
Dimension (mm)	180 (L) x 140 (W) x 390 (H)
Approx. Weight	12 kg



### Set Consist Of:

Model Number	Parts Description	Qty
NL 3000 X / 001 – P 001	Manometer U-Tube in Glass	1no.
NL 3000 X / 001 – P 002	Cell with Perforated Disc & Plunger	1no.
NL 3000 X / 001 – P 003	Rubber Aspirator with Rubber Tube	1no.
NL 3000 X / 001 – P 004	Manometer Liquid 250 ml Bottle	1btl.
NL 3000 X / 001 – P 005	Filter Papers 100 pcs / pack	1pck.
NL 3000 X / 001 – P 006	Panel Board with Stand	1no.
NL 3000 X / 001 – P 007	Standard Reference Cement 114 g to ASTM/SRM/EN to Calibrate the Blaine	1btl.
NL 3000 X / 001 – P 008	Vaseline Grease	1btl.
NL 7034 X / 005 A	Glass Thermometer -10°C to 50°C x 1°C Div.	1no.

\*1 Copy of Manual Instruction