



PLUNGER PENETRATION APPARATUS

Standard: EN 413-2, EN 459-2, EN 1015-4

Consistency of Masonry

The apparatus used for determine the consistency of fresh mortar, lime and masonry cement. Consisting of a steel base with recess to house the test cup and the vertical column holding the penetration plunger assembly with drop height of 100 mm.

Technical Specifications :

Model Number	NL 3011 X / 001
Dimension (mm)	200 (L) x 200 (W) x 700 (H)
Approx. Weight	9 kg

Unit Consist Of:

Model Number	Parts Description	Qty
NL 3011 X / 001 – P 001	Test Cup (80 mm Ø x 70 mm depth)	1no.
NL 3011 X / 001 – P 002	Tamper	1no.
NL 3011 X / 001 – P 003	Metal Sheathed	1no.
NL 3011 X / 001 – P 004	Rigid Stand	1no.
NL 3011 X / 001 – P 005	Penetration Plunger (90 g)	1no.

*1 Copy of Manual Instruction